




America

# CERTIFICATE

No. U8V 101556 0003 Rev. 03

**Holder of Certificate:** **Communications & Power Industries  
Canada Inc.**  
45 River Drive  
Georgetown ON L7G 2J4  
CANADA

**Certification Mark:**



**Product:** **General Medical Devices  
Medical Electrical Component,  
Indico IQ High Frequency X-ray Generator**

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

**Test report no.:** 7169004676-200

**Date,** 2021-12-13

( Peter Keith )



America

# CERTIFICATE

No. U8V 101556 0003 Rev. 03

**Model(s):**

VZW2558FJ8YY, VZW2558FD8YY, VZW2558FC8YY, VZW2558FB8YY, VZW2949FH2-YY

Where:

'F' represents Radiographic or Fluoroscopic X-ray modes and can be R or F  
'YY' represents other hardware and software options not affecting safety and can be 00 to 99 or AA to ZZ.

**Tested according to:**

CAN/CSA-C22.2 No. 60601-2-54:2011  
CAN/CSA-C22.2 No. 60601-1:2014 excluding Clause 11.7, 12.2, 12.3, 15.1 and 17  
ANSI/AAMI ES60601-1:2005/A1:2012-08 excluding Clause 11.7, 12.2, 12.3, 15.1 and 17

**Parameters:**

Rated Input Voltage: 400-480Vac, 3 Phase  
Rated Frequency: 50/60Hz  
Rated Input Current: VZW2558FJ8YY: 4A Nominal, 220-170A Momentary  
VZW2558FD8YY: 4A Nominal, 173-130A Momentary  
VZW2558FC8YY: 3A Nominal, 136-118A Momentary  
VZW2558FB8YY: 3A Nominal, 107-90A Momentary  
VZW2949FH2-YY: 5A Nominal, 65-55A Momentary  
Protection Class: I